## Amendment in technical specification:

## Clause 1(i)(l)

- **Original:** Must include a suitable vacuum pump for microfocus X-ray source to prevent radiation damage of the mirrors and to give long-term stability.
- **Amendment:** If required, must include a suitable vacuum pump for microfocus X-ray source to prevent radiation damage of the mirrors and to give long-term stability.

## **Fig. 1(c)**

- Original: A motorized Z-adjustment with at least 5 mm travel range should be integrated with  $\phi$ -circle.
- **Amendment:** A motorized Z-adjustment with at least 3 mm travel range should be integrated with φ-circle.